Two- and Three-Dimensional Multinuclear Stray-Field Imaging of Rotating Samples with Magic-Angle Spinning (STRAFI-MAS): From Bio to Inorganic Materials

Alan Wong, Dimitris Sakellariou

► To cite this version:

Alan Wong, Dimitris Sakellariou. Two- and Three-Dimensional Multinuclear Stray-Field Imaging of Rotating Samples with Magic-Angle Spinning (STRAFI-MAS): From Bio to Inorganic Materials. *Journal of Magnetic Resonance Imaging*, 2010, 32, pp.418-423. 10.1002/jmri.22242 . cea-00854047

HAL Id: cea-00854047

<https://hal-cea.archives-ouvertes.fr/cea-00854047>

Submitted on 29 Jun 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Original Research

Two- and Three-Dimensional Multinuclear Stray-Field Imaging of Rotating Samples with Magic-Angle Spinning (STRAFI-MAS): From Bio to Inorganic Materials

Alan Wong, PhD and Dimitris Sakellariou, PhD*

Purpose: To revisit and illustrate the potential of a simple and effective multidimensional stray-field imaging technique with magic-angle spinning, known as STRAFI-MAS.

Materials and Methods: STRAFI-MAS images are acquired with a standard NMR magnet and a traditional magic-angle sample spinning (MAS) probe. The stray-field gradients are achieved by placing the MAS probe, along the z -direction, at a distance from the center of the magnet. No pulsed-field gradients are applied. The multidimensional spatial encoding is carried out by synchronizing the radiofrequency pulses with the sample MAS rotation.

Results: Two-dimensional (2D) and 3D multinuclear images of various phantoms, including a tibia bone and silicon carbide, are recorded. Images of inorganic solids containing quadrupolar nuclei, ^{23}Na and ^{27}Al , are also explored for the first time by STRAFI-MAS.

Conclusion: We have demonstrated that STRAFI-MAS is a simple and user-friendly technique for multidimensional imaging without the need of imaging equipment. With the current advancements in NMR and MRI methodologies, STRAFI-MAS is expected to be further developed and improved. We anticipate that STRAFI-MAS can spark a wide spectrum of interest, from material to bio science, where can benefit from high-resolution images.

Key Words: stray-field imaging; magic-angle rotating sample; multidimensional imaging; multinuclear imaging; large field gradient

J. Magn. Reson. Imaging 2010;32:418–423.

© 2010 Wiley-Liss, Inc.

MRI IS AN IMAGING technique used primarily in medicine to produce high quality images for medical analysis. The technique is based on the principles of nuclear MR (NMR), a spectroscopy used by scientists to obtain microscopic chemical and physical information of liquid and solid phase materials. MRI was first demonstrated by a back reconstruction projection of ^1H NMR frequency in the presence of a magnetic field gradient (1,2). This is, in fact, still the fundamental basis for today's MRI experiments. Field gradients along the x , y , and z direction are necessary for encoding multidimensional k -space (raw image space) of a 3D sample-frame. Generally, the pulsed gradient strengths of no greater than 3 T/m are used in most clinical and research facilities. There is a constant strive to apply higher gradients because the spatial resolution is directly dependent to the applied field gradient: the stronger the gradient, the higher the resolution. Samoilenko et al (3) was first to propose the use of a large stray-field gradient, up to 100 T/m, from the fringe field of a superconducting NMR magnet. However, stray-field imaging (STRAFI) is a 1D technique in nature, due to the fact that it involves a single directional gradient along the z -axis of the laboratory frame. To overcome this handicap, there are generally two approaches toward multidimensional STRAFI. The first is by adding external pulsed-field gradients (along the x and y direction) for multidimensional k -space encoding (4); however, this would result in an uneven spatial resolution, high along the z direction and low along the x and y . The second approach is by mechanically reorienting the sample with respect to a single-directional fixed gradient (5), in such that the k -space encoding can be carried by a time consuming back reconstruction projection. In both approaches, specialized equipments are required and are not easily accessible.

High spatial resolution in solid samples requires additional line-narrowing techniques. For example, the first-order anisotropic interactions, such as chemical shift, dipolar coupling and magnetic susceptibility, have an angular dependence proportional to $(3\cos^2\Theta - 1)$, where Θ is the angle with respect to the external magnet field B_0 . These can be readily

UMR 3299 CEA Saclay, DSM/IRAMIS/CNRS-SIS2M/LSDRM, F-91191, Gif-sur-Yvette Cedex, France.

Contract grant sponsor: European Research Council under the European Community's Seventh Framework Programme; Contract grant number: FP7/2007-2013; Contract grant sponsor: ERC; Contract grant number: 205119.

*Address reprint requests to: D.S., CEA Saclay, DSM/IRAMIS/SIS2M/LSDRM, F-91191, Gif-sur-Yvette Cedex, France.
E-mail: dimitrios.sakellariou@cea.fr

Received February 8, 2010; Accepted April 29, 2010.

DOI 10.1002/jmri.22242

Published online in Wiley InterScience (www.interscience.wiley.com).

averaged by rotating the sample rapidly at 54.74° about B_0 . This line-narrowing technique is called magic-angle sample spinning (MAS) (6,7), and is now a routine experiment in NMR spectroscopy for solid materials. The concept of MAS imaging have been proposed by Wind and Yannoni (8). But the first MAS imaging experiment was implemented by Cory et al (9,10), where the gradient coils were built so that the gradients rotate synchronously with the MAS rotor, resulting in gradient components fixed with the sample body frame. An alternative to synchronize the gradients with the MAS rotor (i.e., the sample body frame) is by simulating the effective rotating gradients using external gradient pulses along the x , y , and z laboratory frame. Pampel (11) has demonstrated the latter approach by inserting a standard 4-mm high-resolution MAS probe into a standard pulsed-field gradient system (Bruker Micro2.5). Although the above two MAS imaging techniques promise access to high resolution imaging, they both require technically demanding hardware.

In 2005, Baltisberger et al (12) have proposed an ingenious approach for acquiring multidimensional images by combining the stray-field gradient and the line-narrowing MAS technique, where MAS is also used as a tool for sample reorientation. This technique is called STRAFI-MAS. Figure 1A illustrates the principle of STRAFI-MAS, as the sample under a continuous rotation (ν_R) at the magic-angle (54.74°) about B_0 , the body-frame axes (x , y , and z) rotates and changes its orientations successively every third ($T_r/3 = 120^\circ$) of a rotation. In other words, after a 120° rotation of the sample holder (known as the "rotor") at the magic-angle, the z -body axis becomes x , and x becomes y , and y becomes z . Thereby, all three body-frame axes have successive aligned with the fixed stray-field z -gradient. As a result, the spin magnetization density can be encoded at each body-axis during the three periods, t_1 , t_2 , and t_3 . The encoding can be carried by a modified magic-angle turning (MAT) pulse sequence (13). A 3D experiment (Fig. 1B) begins with two 90° pulses which are centered at every one-third of the rotor period, creating a balanced effective gradient orientation during the periods, t_1 , t_2 , and t_3 . The two pulses excite and store the evolved magnetization, at each of the three body orientations with respect to the fixed gradient for incremented times Δt_1 , Δt_2 , and Δt_3 , by simply tipping (first 90°) the magnetization to the transverse plane, and projecting (second 90°) back to the longitudinal z -axis of laboratory frame at every one-third rotor period. The last 180° pulse is to eliminate image distortions, in the t_3 period, from the probe/receiver dead-time. In this experiment, the spatial k -space encoding is done by frequency encoding with a single fixed stray-field gradient along the z -axis. Similarly, a 2D experiment can be performed by encoding the frequency during two of the three sample orientations. A 2D k -space encoding scheme is illustrated in Figure 1C. Additional CPMG (Carr-Purcell-Meiboom-Gill) (14) train pulses could be applied as an option for signal enhancement. The spatial resolution and/or the sample size is limited by the excitation bandwidth generates from the radiofrequency (rf)

Figure 1. **A:** Diagram showing the sample body-frame axes, x , y , and z , undergoing a change of its orientation after a successive 120° rotation along the magic-angle axis: $z \rightarrow y \rightarrow x$. **B:** STRAFI-MAS pulse experiments for encoding the k -space along the x , y , and z body-axes to give rise to 2D and 3D spin density images. The coherence pathways of the experiment were selected as described in Baltisberger et al (12). T_r is one rotor period of a MAS rotation. The inserted circles indicate the sample rotational phase. τ_{rp} is a rotor (i.e., sample) projection period for setting the rotor projection angle Θ_{rp} of the image: $(1/\nu_R) \times (\Theta_{rp}/360^\circ)$, where ν_R is a MAS frequency. **C:** The 2D k -space sampling scheme, where Δt is the dwell time in t_1 and t_2 .

power of a probehead. However, with the advancements in NMR technology, high rf-power MAS probe is readily available. Baltisberger et al (12) have demonstrated the experiment and presented 2D ^1H images.

To the best of our knowledge, there are no further developments and applications of STRAFI-MAS have been reported. Here, we revisit the STRAFI-MAS imaging technique with 2D and 3D image acquisitions, and explore with different spin- $1/2$ and spin $> 1/2$ nuclei to illustrate its potential application to a wide range of materials.

MATERIALS AND METHODS

The STRAFI-MAS images of water phantoms and a natural abundance silicon carbide (SiC, purchased from Aldrich, France) were carried out on a 11.75T widebore (89 mm) Bruker 500 UltraShield™ magnet with a Avance III 500 spectrometer, and a standard Bruker 7 mm HX-CPMAS probe. The STRAFI-MAS images of a tibia bone and a $\text{Na}_2\text{SiO}_3/\text{Al}(\text{OH})_3$ (Aldrich, France) mixed phantom were performed on a 7.05T widebore (89 mm) Oxford magnet with a Avance 300 spectrometer, and a Bruker 4 mm HXY-CPMAS probe. The field gradients (0.05–1.0 T/m) used in this study were achieved by simply positioning the MAS probe at an offset, along the bore of a NMR magnet, from its standard position (sweep-spot of the magnet). The gradient strengths were carefully measured from the ^1H linewidth of a known dimension water phantom. It should be pointed out that the gradients applied here are comparable to those offered from a standard commercial pulsed-field gradient system, up to 3 T/m. The probe's magic-angle, 54.74° , was carefully calibrated with ^{89}Br NMR of KBr. The MAS frequency was manually adjusted using a standard Bruker pneumatic controller. Spinning frequencies of less than 200 Hz with stability less than ± 1 Hz were achieved with a custom-made turbine-less Kel-F rotor cap. All multidimensional images were recorded with States phase cycling (15) and processed with an exponential apodization on the t_1 and t_2 time-domain data to improve the signal-to-noise ratio. Zero-filling was also applied to improve the pixel definition quality. Further experimental and processing details are provided in the figure captions.

RESULTS AND DISCUSSION

^1H Images of Water Phantoms

Figure 2A shows a 2D ^1H image of a Kel-F insert having two chambers filled with water inside a 7-mm MAS rotor, with a resolution about $185\ \mu\text{m} \times 730\ \mu\text{m}$. The tilted image (pointing at the magic-angle) can be readily sheared by a simple 2D rotational matrix to a favorable representation. Although the image is acquired with a relatively low stray-field z-gradient (< 0.1 T/m), it shows the physical features of the phantom: two water chambers with an unsmooth surface all around the wall. The 2D image is carried out under a slow continuous magic-angle rotation, having a frequency of 105 Hz with stability of less than ± 1 Hz. The slow and stable rotation reduces the image distortion arising from the change in the gradient orientation during t_1 and t_2 (12). Moreover, short echo period ($\Delta = 5\ \mu\text{s}$) is used to minimize the distortion

Figure 2. A: Unsheared and sheared ^1H 2D image of a water phantom of two-chamber 7-mm Kel-F insert. The phantom is displayed on the right. Experimental details: probe offset 5.8 cm; ^1H frequency at 499.12 MHz; z-gradient ~ 0.05 T/m; 105 ± 1 Hz MAS; $2.5\ \mu\text{s}$ 90° pulse width (^1H -rf of 100 kHz); $0\ \mu\text{s}$ τ_{rp} ; $5\ \mu\text{s}$ Δ ; 0.25 s recycle delay; 8 scans; 512 and 128 real data points were collected for a 200,000 Hz spectral width in F_2 and F_1 , respectively. The total acquisition time is 9 min. The image was processed with an exponential apodization of 200 Hz in t_2 and $t_1 \cdot t_2$ and t_1 data were zero-filled to 2048 points to increase the pixel resolution; however, the actual image resolution ca. $185\ \mu\text{m} \times 730\ \mu\text{m}$ is unaffected. **B:** Water phantom of a 4-mm round bottle flask inside a 7-mm MAS rotor. Experimental details: probe offset 7.9 cm; ^1H frequency at 498.88 MHz; z-gradient ~ 0.62 T/m; 111 ± 1 Hz MAS; $2.5\ \mu\text{s}$ 90° pulse width (^1H -rf of 100 kHz); [for 1D projection] 0 Hz MAS (i.e., static sample); acquired with a full-echo; 5 s recycle delay; 16 scans; [for 2D projection] 0.5 s recycle delay; 32 scans; $0\ \mu\text{s}$ τ_{rp} ; $5\ \mu\text{s}$ Δ ; 1024 and 50 real data points were collected for a 500,000 Hz spectral width in F_2 and F_1 , respectively. Acquisition time is 30 min. The image was processed with an exponential apodization of 500 Hz in t_2 and t_1 , and zero-filled to 2048 and 512 points, respectively. The image resolution is about $19\ \mu\text{m} \times 380\ \mu\text{m}$; [for 3D projection] 0.7 s recycle delay; 32 scans; $5\ \mu\text{s}$ Δ ; 1024, 26 and 50 real data points were collected for a 500,000 Hz spectral width in F_3 , F_2 , and F_1 , respectively. The acquisition time is 32 hr. The image was processed with an exponential apodization of 1000 Hz in t_3 , t_2 , and t_1 , and zero-filled to 2048, 64 and 128 points, respectively. The image resolution is ca. $19\ \mu\text{m} \times 743\ \mu\text{m} \times 379\ \mu\text{m}$.

arising from the slight difference in the gradient orientation during each Δ period (12). Thus, slow and stable MAS conditions with short Δ minimize the

Figure 3. A photograph of a hollow tibia bone (left) and its sheared ^1H images with two different rotor projection angles θ_{rp} , 0° and 90° (right). Experimental details: probe offset 9.7 cm; ^1H frequency at 298.12 MHz; z-gradient ~ 0.9 T/m; 1050 ± 5 Hz MAS; 2 μs 90° pulse width (^1H -rf of 125 kHz); 0 and 238 μs τ_{rp} , which correspond to θ_{rp} 0° and 90° , respectively; 5 μs Δ ; 1 s recycle delay; 64 scans, 512 and 64 real data points were collected for a 1,000,000 Hz spectral width in F2 and F1. The acquisition time is 2.3 hr. The image was processed with an exponential apodization of 3000 Hz in t_2 and t_1 , and zero-filled to 1024 and 512 points, respectively. The image resolution is ca. $51 \mu\text{m} \times 408 \mu\text{m}$.

imperfection of gradient orientation, and provide good quality images. Similarly, Figure 2B shows 1D, 2D, and 3D images of water inside a 4-mm round-bottle flask. The images were acquired with a gradient of ~ 0.62 T/m. The resultant resolutions are estimated to be $19 \mu\text{m} \times 379 \mu\text{m}$ for 2D, and $19 \mu\text{m} \times 743 \mu\text{m} \times 379 \mu\text{m}$ for 3D. The images are acquired under a slow MAS (spinning frequency of 111 Hz) with a short Δ period (5 μs) to eliminate distortions, and are in good agreement with the water phantom. Note that the 3D image appears to be tilted, and is slightly distorted around the round bottle region. The latter is contributed from the fact that the signal is truncated during the t_2 period, and results in a poor pixel resolution (743 μm) along the t_2 dimension. An improved image could be acquired with longer acquisition time. Nevertheless, the 3D image reveals the round bottle feature quite well, and it represents the first three-dimensional image acquired by STRAFI-MAS.

^1H Image of a Tibia Bone

The 2D STRAFI-MAS images of a small hollow tibia bone ($\sim 6 \text{ mm} \times \sim 1.5 \text{ mm}$) are carried out at 7.05T with a stray-field gradient of ~ 0.9 T/m, and are shown in Figure 3. The resolution is approximately $51 \mu\text{m} \times 408 \mu\text{m}$. The probe was positioned 9.7 cm below the center of the magnet. Although the image's feature does not agree well with the tibia bone shown in the photograph on the left, it does display the high ^1H density regions at the two long edges, and the low density at the center (hollow) region. Moreover, the two images with different rotor (i.e., sample) projection angles ($\theta_{rp} = 360^\circ \times \nu_R \times \tau_{rp}$, where ν_R is a MAS

frequency and τ_{rp} is the rotor projection period used in the experiment), 0° and 90° , do correspond modestly with the view projection of the bone. For example, the image with a 90° rotor projection corresponds to a mirror reflection of the photograph; whereas, the image with 0° rotor projection corresponds to a 90° turn of the photographic picture. The visible distortions in the images arise from the fact that the experiments were carried out in an unstable MAS rotation, 1050 Hz with a constant fluctuation of ± 5 Hz. A better image quality could be obtained with a slower and a more stable MAS rotation. This could be achieved by using a larger MAS rotor with a turbine-less cap. It is interesting to note that Tritt-Goc et al (16) have acquired high spatial 3D resolution images ($60 \times 60 \times 60 \mu\text{m}$) of a human trabecular bone at 7T to investigate a bone disease called osteoporosis. In the study, a field gradient of 1 T/m was used along the x , y , and z direction, which is comparable to what we applied here for the tibia bone.

^{29}Si Image of Silicon Carbide

Silicon is a widely used element in many applications, ranging from high endurance materials (i.e., automobile components, bearings, and abrasives) to electronic and nuclear energy applications (i.e., radioactive storage components, emitting diodes, detectors, and fuel). Thereby ^{29}Si MAS experiments are vastly applied to study the atomic structure around the silicon sites (17). Figure 4 shows a ^{29}Si image of an industrial importance silicon material, SiC powder, packed without any effort to generate a featured object, at the center region of a 7-mm rotor. It was acquired with a gradient of ~ 0.45 T/m, and a stable MAS rotation frequency of 1800 Hz. The image resolution is approximately $56 \mu\text{m} \times 521 \mu\text{m}$. The image dimensions correspond well with those of the phantom. The distortion of nonparallel edges is attributed to a combination of the fast MAS and the inherently low sensitive ^{29}Si nucleus. The low ^{29}Si natural

Figure 4. ^{29}Si image of natural abundance SiC powder tightly packed at the center of a 7-mm rotor. Experimental details: probe offset 7 cm; ^{29}Si frequency at 99.16 MHz; z-gradient ~ 0.45 T/m; 1800 ± 1 Hz MAS; 3.4 μs 90° pulse width (^{29}Si -rf of 74 kHz); 0 μs τ_{rp} ; 5 μs Δ ; 330 s recycle delay; 32 scans, 150 and 16 real data points were collected for a 50000 Hz spectral width in F_2 and F_1 . The acquisition time is 39 hr. The image was processed with an exponential apodization of 1000 Hz in t_2 and t_1 , and zero-filled to 2048 and 512 points, respectively. The image resolution is ca. $56 \mu\text{m} \times 521 \mu\text{m}$.

abundance (4.7%), and the long longitudinal relaxation time T_2 (1.9 min) for ^{29}SiC (18) result in a very time consuming experiment. Fortunately, in both crystalline and amorphous natural abundance samples, the intrinsic transverse relaxation time T_2 of the ^{29}Si nuclei can be much longer than T_2^* rendering the possibility of observing successive Hahn echoes in a CPMG train, and enhancing the sensitivity during the acquisition (19,20). Larsen and Farnan (19) has demonstrated that 1D ^{29}Si NMR signal enhancement with ^{29}Si MAS CPMG on silicon glasses. Similarly, the ^{29}Si signal efficiency of implementing CPMG into a high-resolution 2D MAT experiment has also been demonstrated (20). Thereby in principle, a signal enhanced STRAFI-MAS image could be readily achieved by coupling the CPMG during the signal acquisition period, see Figure 1B. ^{29}Si STRAFI-MAS with CPMG is a direction that we plan on pursuing and demonstrate in a near future.

^{23}Na And ^{27}Al Images of Inorganic Salts

To investigate the extension of STRAFI-MAS to quadrupolar nuclei ($\text{spin} > 1/2$), we chose to image with ^{23}Na ($I = 3/2$) and ^{27}Al ($I = 5/2$), which are abundance elements in materials, and are commonly characterized by MAS spectroscopy. Due to the additional quadrupolar interaction in the quadrupolar nuclei, it often resulted in fast spin-spin transverse relaxation and broad resonance, hindering high resolution imaging. Despite this, ^{23}Na MRI is still a good diagnostic technique for localizing and quantifying the sodium concentration in materials such as foods (21), soft tissues (22), bone cartilages (23), and cements (24). Most of today's ^{23}Na imaging is done on nonrigid samples, or highly mobile sodium ions, where the quadrupolar interaction is expected to be small. For rigid samples, the second-order quadrupolar interaction can become sufficient and hampering for ^{23}Na imaging. Even under MAS, the resonance lines may show certain frequency distributions arising from the second-order anisotropic interaction yielding quadrupolar parameters: $\nu_Q = [3C_Q/2I(2I-1)](1+\eta^2/3)^{0.5}$, where C_Q and η are quadrupolar coupling constant and asymmetry parameter, respectively. Figure 5 shows ^{23}Na and ^{27}Al STRAFI-MAS images of a solid phantom consists of $\text{Al}(\text{OH})_3$ and Na_2SiO_3 powder layers each with 1 – 2 mm thickness. Under a fast but stable 1500 Hz MAS, we were able to acquire ^{23}Na and ^{27}Al images with full-echo acquisition (one rotor period for Δ) to eliminate image distortion from the imperfection of magnetization refocusing during the t_2 period. However, the distortions around the edges are mainly attributed to the second-order quadrupolar interaction. In the case, the ^{27}Al image is slightly more distorted compared to ^{23}Na . This is due to the fact that $\text{Al}(\text{OH})_3$ consists of two ^{27}Al sites with different quadrupolar frequency ν_Q , 0.24 and 0.67 MHz (25); whereas, Na_2SiO_3 has only one ^{23}Na site with ν_Q of 0.79 MHz (26). Despite the distortion, the ^{23}Na and ^{27}Al images shown in Figure 5 clearly display the individual layers containing sodium in Na_2SiO_3 and aluminum in $\text{Al}(\text{OH})_3$. In principle, a magic-angle turning

Figure 5. (Left) A $\text{Na}_2\text{SiO}_3/\text{Al}(\text{OH})_3$ phantom with the number (in millimeters) indicating the estimated thickness of each layer. (Right) ^{23}Na and ^{27}Al STRAFI-MAS images. Experimental details: probe offset 8.6 cm; ^{23}Na and ^{27}Al frequency at 79.14 and 77.95 MHz, respectively; z-gradient ~ 0.47 T/m; 1500 ± 1 Hz MAS; [for ^{23}Na] $1.13 \mu\text{s}$ solid- 90° pulse width (^{23}Na -rf of 111 kHz); $0 \mu\text{s}$ τ_{rp} ; acquired with a full-echo; 3-s recycle delay; 64 scans; 512 and 46 real data points were collected for a 1,000,000 Hz spectral width in F_2 and F_1 . The acquisition time is 4.8 hr. The image was processed with an exponential function of 1200 Hz in t_2 and t_1 , and zero-filled to 4096 and 512 points, respectively; [for ^{27}Al] $1.25 \mu\text{s}$ solid- 90° pulse width (^{27}Al -rf of 67 kHz); $0 \mu\text{s}$ τ_{rp} ; acquired with a full-echo; 0.1 s recycle delay; 2024 scans; 512 and 28 real data points were collected for a 1,000,000 Hz spectral width in F_2 and F_1 . The acquisition time is 3 hr. The image was processed with an exponential apodization of 1500 Hz in t_2 and t_1 , and zero-filled to 2048 and 512 points, respectively.

experiment for quadrupolar nuclei (QMAT) (27) could be incorporated in the STRAFI-MAS experiment to minimize the distortion by suppressing the inherent MAS sidebands. Alternatively, a larger magnetic field gradient, such that it dominates the nuclear quadrupolar interaction, could be used to improve the spatial resolution and suppress the distortion from the quadrupolar interaction. Notice that very large stray-field gradients are easily accessible away from the center of any NMR magnet.

The continuous sample rotation is a main factor contributing to the image distortion in STRAFI-MAS (12). It arises from the imperfection of gradient orientation in t_1 , t_2 , and t_3 evaluation periods, and the sample vibration during the course of sample rotation. We minimized these effects by means of slow and stable MAS sample rotation, and showed good quality 2D images of water phantoms in Figure 2. Alternatively, the distortion from the sample rotation could be “completely” eliminated by using a discrete sudden 120° rotation (rather than a continuous rotation, MAS), about an axis oriented at the magic-angle to B_0 , between the t_1 , t_2 , and t_3 periods. This, in fact, is the principle of a magic-angle hopping (MAH) experiment (28) for obtaining high-resolution solid-state NMR spectroscopy. In principle, the STRAFI-MAS experiment could be served as a protocol for incorporating

MAH with STRAFI as a new and improved approach for multidimensional NMR imaging.

In conclusion, this study demonstrates a potentially useful technique, STRAFI-MAS, for multidimensional and multinuclear imaging. It utilizes the easily accessible large gradient from the stray-field of a NMR magnet, and the slow sample reorientations around the magic-angle. In addition, the experiment is user-friendly and simple to set up. It only requires a standard MAS probe capable of stable slow spinning. Although at this point, STRAFI-MAS is a primitive multidimensional imaging technique, but there is plenty of room for further developments. For example, adding CPMG echo pulses would enhance signal sensitivity; and implementing T_1 , $T_{1\rho}$, or multiquantum filter pulses would provide contrasting images, which are an enormous advantage for MRI over other imaging techniques. With the accessible large gradient and the simplicity of the experiment, we anticipate that STRAFI-MAS is a promising imaging technique for a wide range of microscopic studies, including micro- or maybe even nano-scale imaging by adopting the recent developments of micro-sized slow MAS NMR spectroscopy (29,30). One highly possible interest would be cell-imaging, where it can benefit from the large stray-field field gradient for ultra-high resolution.

ACKNOWLEDGMENTS

We thank Dr. Sabine Hediger (Lyon, France) for the helpful discussions on the STRAFI-MAS experiment, and Dr. Pedro Aguiar and Mr. Angelo Guiga (Gif-sur-Yvette, France) for their contributions on insert fabrication. We are also grateful to Dr. Sara Laurencin Dalicieux (France) for providing the mouse tibia bone. We also thank the referees for their valuable comments.

REFERENCES

- Lauterbur PC. Image formation by induced local interactions: examples employing nuclear magnetic resonance. *Nature* 1973; 242:190–191.
- Mansfield P, Grannell PK. NMR 'diffraction' in solids? *J Phys C Solid State Phys* 1973;6:L422–L426.
- Samoilenko AA, Artemov DY, Sibeldina LA. Formation of sensitive layer in experiments on NMR subsurface imaging of solids. *JETP Letters* 1988;47:417–419.
- Godward J, Ciampi E, Cifelli M, McDonald PJ. Multidimensional imaging using combined stray field and pulsed gradients. *J Magn Reson* 2002;155:92–99.
- Samoilenko AA, Zick K. Stray field imaging of solids. *Bruker Report* 1990;1:40–41.
- Andrew ER, Bradbury A, Eades RG. Nuclear magnetic resonance spectra from a crystal rotated at high speed. *Nature* 1958;182:1659.
- Lowe I. Free induction decays of rotating solids. *Phys Rev Lett* 1959;2:285–287.
- Wind RA, Yannoni CS. Spin imaging in solids using synchronously rotating field gradients and samples. US patent 4,301,410, 17 November 1981.
- Cory DG, Reichwein AM, Van OS-JWM, Veeman WS. NMR images of rigid solids. *Chem Phys Lett* 1988;143:467–470.
- Cory DG, Van OS-JWM, Veeman WS. NMR images of rotating solids. *J Magn Reson* 1988;76:543–547.
- Pampel A, Engeike F, Groß D, Oerther T, Zick K. Pulse-field-gradient NMR with magic-angle spinning: diffusion studies and NMR microscopy of heterogeneous materials. *Bruker Spin Report*, 2006;157/158:26–30.
- Baltisberger JH, Hediger S, Emsley L. Multi-dimensional magnetic resonance imaging in a stray magnetic field. *J Magn Reson* 2005;172:79–84.
- Gan Z. High-resolution chemical shift and chemical shift anisotropy correlation in solids using slow magic-angle spinning. *J Am Chem Soc* 1992;114:8307–8309.
- Waugh JS. Sensitivity in Fourier transforms NMR spectroscopy of slowly relaxing systems. *J Mol Spectrosc* 1970;35:298–305.
- States DJ, Haberkorn RA, Ruben DJ. A two-dimensional nuclear overhauser experiment with pure absorption phase in four quadrants. *J Magn Reson* 1982;48:286–292.
- Tritt-Goc J, Piślewski N, Hachulska AK, Chmielewski D, Górecki A, Kolodziejewski W. Proton magnetic resonance microimaging of human trabecular bone. *Solid State Nucl Magn Reson* 1999;15: 91–98.
- Farnan I, Cho H, Weber WJ. Quantification of actinide alpha-radiation damage in minerals and ceramics. *Nature* 2007;445: 190–193.
- Leonova E, Grins J, Shariatgorji M, Ilag LL, Edén M. Solid-state NMR investigations of Si-17 and N-15 enriched silicon nitride. *Solid State Nucl Magn Reson* 2009;36:11–18.
- Larsen FH, Farnan I. ^{29}Si and ^{17}O (Q)CPMG-MAS solid-state NMR experiments as an optimum approach for half-integer nuclei having long T_1 relaxation times. *Chem Phys Lett* 2002;357: 403–408.
- Sakellariou D, Jacquinet J-F, Charpentier T. 2D correlation spectra of isotropic and anisotropic ^{29}Si chemical shifts in crystalline and amorphous natural abundance materials under very slow sample rotation. *Chem Phys Lett* 2005;411:171–174.
- Veliyulin E, Egelanddal B, Marica F, Balcom BJ. Quantitative ^{23}Na magnetic resonance imaging of model foods. *J Agric Food Chem* 2009;57:4091–4095.
- Maril N, Rosen Y, Reynolds GH, Ivanishev A, Ngo L, Lenkinski RE. Sodium MRI of the human kidney at 3 Tesla. *Magn Reson Med* 2006;56:1229–1234.
- Borthakur A, Mellon E, Niyogi S, Witschey W, Kneeland JB, Reddy R. Sodium and T1rho MRI for molecular and diagnostic imaging of articular cartilage. *NMR Biomed* 2006;19:781–821.
- Cano FdJ, Bremner TW, McGregor RP, Balcom BJ. Magnetic resonance imaging of ^1H , ^{23}Na , and ^{35}Cl penetration in Portland cement mortar. *Cem Concr Res* 2002;32:1067–1070.
- Faucon P, Charpentier T, Bertrandie D, Nonat A, Virlet J, Petit JC. Characterization of calcium aluminate hydrates and related hydrates of cement pastes by ^{27}Al MQ-MAS NMR. *Inorg Chem* 1998;37:3726–3733.
- Koller H, Engelhardt G, Kentgens A, Sauer J. ^{23}Na NMR spectroscopy of solids: interpretation of quadrupole interaction parameters and chemical shifts. *J Phys Chem* 1994;98:1544–1551.
- Nakai T, Kuwahara D. Spinning sideband summation for quadrupole MAS NMR spectra using a magic-angle turning technique. *Chem Phys Lett* 1996;249:205–209.
- Bax AD, Szeverenyi NM, Maciel GE. Correlation of isotropic shifts and chemical shift anisotropies by two-dimensional Fourier-transform magic-angle hopping NMR spectroscopy. *J Magn Reson* 1983;52:147–152.
- Sakellariou D, Goff GL, Jacquinet J-F. High-resolution, high-sensitivity NMR of nanolitre anisotropic samples by coil spinning. *Nature* 2007;447:694–697.
- Wong A, Augiar PM, Sakellariou D. Slow magic-angle coil spinning: a high-sensitivity and high-resolution NMR strategy for microscopic biological specimens. *Magn Reson Med* 2010;63: 269–274.